



ITC Engineering Services

Certificate of Compliance

Rev. 2

Application of Council Directive: 2004/108/EC - EMC Directive)

Standards to which Conformity is Declared:
EN 61326-1:2013 including: EN 55011:2009, Group 1, Class A;
EN 61000-3-2:2006/A2:2010; EN 61000-3-3:2008;
EN 61000-4-2:2009; EN 61000-4-3:2006/A2:2010;
EN 61000-4-4:2012; EN 61000-4-5:2006; EN 61000-4-6:2009;
EN 61000-4-8:2010; EN 61000-4-11:2004

Applicant:

Power Standards Lab

Address:

2020 Challenger Drive, Ste. 100

Alameda, CA 94501

Tel: 510.522.4400 Fax: 510.522.4455

Product Tested:

High-Speed Power Analyzer

Models Tested:

PQube 3

Tested By:

ITC Engineering Services, Inc.

9959 Calaveras Road, PO Box 543

Sunol, California 94586-0543

Tel: 925.862.2944 Fax: 925.862.9013

Email: info@itcemc.com

http://www.itcemc.com

Date of Issue:

March 28, 2014

Report Numbers:

20140127-01

I, the undersigned hereby declare that the models listed above were tested and deemed to conform to the Directives and Standards listed above.

Certified By:

Date:

March 28, 2014

Mr. Michael Gbadebo, PE
(California License # 11303)
Chief Engineer/Principal Consultant



ISO/IEC 17025:2005 Accredited Laboratory

