



Power Standards Lab
 980 Atlantic Avenue
 Alameda, CA 94501 USA
 TEL ++1-510-522-4400
 FAX ++1-510-522-4455
 www.PowerStandards.com

Calibration Certificate No. P3001143-05072015

PQube 3 S/N P3001143

Calibration date: 07 May 2015
Calibration time: 4:15 PM
Calibration location: PSL – Alameda, CA
Temperature: 24°C
Relative humidity: 47%
Calibration software: 1.3.7
Calibrated by: Chuck Lopez

Reference or Source	Serial Number	NIST (or NIST-equivalent) Trace Path	Due Date	Equip. Used
Leeds & Northrup 4221-B 0.1-Ω Standard Resistor	1709567	NIST 817/271591-05	—	
Fluke 731B DC Reference Standard	4455007	NIST 817/271591-05	—	
HP 3458A Digital Multimeter	2823A14288	Simco Cert. No. 6936554	05/06/16	✓
National Instruments USB-9269 Analog Output Module	1876CF6	N/R	N/A	✓
Trek 2210 High-Voltage Amplifier	241	N/R	N/A	✓
Chroma 61601 Programmable AC Source	616010002046	N/R	N/A	✓
HP 33120A Function Generator	US34000902	N/R	N/A	
Fluke 5220A Transconductance Amplifier	2700009	N/R	N/A	

Parameter	PQube 3 Reading	Reference Reading	Reference Uncertainty X% rdg + X% FS	PQube 3 Error X% rdg + X% FS	Published Specification X% rdg + X% FS	Factory Pass/Fail X% rdg + X% FS	Pass/Fail	Notes
L1-N	120.06 V _{rms}	120.04 V _{rms}	± 0.009%	0.00%	± 0.05%	± 0.025%	Pass	50 Hz, FS = 600 V _{rms}
L1-N	230.08 V _{rms}	230.06 V _{rms}	± 0.014%	0.00%	± 0.05%	± 0.025%	Pass	50 Hz, FS = 600 V _{rms}
L1-N	600.30 V _{rms}	600.13 V _{rms}	± 0.022%	0.01%	± 0.05%	± 0.025%	Pass	50 Hz, FS = 600 V _{rms}
L2-N	120.06 V _{rms}	120.04 V _{rms}	± 0.009%	0.00%	± 0.05%	± 0.025%	Pass	50 Hz, FS = 600 V _{rms}
L2-N	230.08 V _{rms}	230.06 V _{rms}	± 0.014%	0.00%	± 0.05%	± 0.025%	Pass	50 Hz, FS = 600 V _{rms}
L2-N	600.27 V _{rms}	600.13 V _{rms}	± 0.022%	0.01%	± 0.05%	± 0.025%	Pass	50 Hz, FS = 600 V _{rms}
L3-N	120.06 V _{rms}	120.04 V _{rms}	± 0.009%	0.00%	± 0.05%	± 0.025%	Pass	50 Hz, FS = 600 V _{rms}
L3-N	230.08 V _{rms}	230.06 V _{rms}	± 0.014%	0.00%	± 0.05%	± 0.025%	Pass	50 Hz, FS = 600 V _{rms}
L3-N	600.27 V _{rms}	600.13 V _{rms}	± 0.022%	0.01%	± 0.05%	± 0.025%	Pass	50 Hz, FS = 600 V _{rms}
N-Earth	120.06 V _{rms}	120.06 V _{rms}	± 0.009%	0.00%	± 0.05%	± 0.025%	Pass	50 Hz, FS = 600 V _{rms}
N-Earth	230.12 V _{rms}	230.11 V _{rms}	± 0.014%	0.00%	± 0.05%	± 0.025%	Pass	50 Hz, FS = 600 V _{rms}
N-Earth	600.30 V _{rms}	600.25 V _{rms}	± 0.022%	0.00%	± 0.05%	± 0.025%	Pass	50 Hz, FS = 600 V _{rms}
L1-N	120.03 V _{rms}	120.00 V _{rms}	± 0.009%	0.00%	± 0.05%	± 0.025%	Pass	60 Hz, FS = 600 V _{rms}
L1-N	230.01 V _{rms}	230.00 V _{rms}	± 0.014%	0.00%	± 0.05%	± 0.025%	Pass	60 Hz, FS = 600 V _{rms}
L1-N	600.09 V _{rms}	599.96 V _{rms}	± 0.022%	0.01%	± 0.05%	± 0.025%	Pass	60 Hz, FS = 600 V _{rms}
L2-N	120.03 V _{rms}	120.00 V _{rms}	± 0.009%	0.00%	± 0.05%	± 0.025%	Pass	60 Hz, FS = 600 V _{rms}
L2-N	230.01 V _{rms}	230.00 V _{rms}	± 0.014%	0.00%	± 0.05%	± 0.025%	Pass	60 Hz, FS = 600 V _{rms}
L2-N	600.09 V _{rms}	599.96 V _{rms}	± 0.022%	0.01%	± 0.05%	± 0.025%	Pass	60 Hz, FS = 600 V _{rms}
L3-N	120.03 V _{rms}	120.00 V _{rms}	± 0.009%	0.00%	± 0.05%	± 0.025%	Pass	60 Hz, FS = 600 V _{rms}
L3-N	230.01 V _{rms}	230.00 V _{rms}	± 0.014%	0.00%	± 0.05%	± 0.025%	Pass	60 Hz, FS = 600 V _{rms}
L3-N	600.09 V _{rms}	599.96 V _{rms}	± 0.022%	0.01%	± 0.05%	± 0.025%	Pass	60 Hz, FS = 600 V _{rms}
N-Earth	120.03 V _{rms}	120.03 V _{rms}	± 0.009%	0.00%	± 0.05%	± 0.025%	Pass	60 Hz, FS = 600 V _{rms}
N-Earth	230.05 V _{rms}	230.06 V _{rms}	± 0.014%	0.00%	± 0.05%	± 0.025%	Pass	60 Hz, FS = 600 V _{rms}
N-Earth	600.16 V _{rms}	600.12 V _{rms}	± 0.022%	0.00%	± 0.05%	± 0.025%	Pass	60 Hz, FS = 600 V _{rms}
AN1-Earth High Range	-70.136 V	-70.135 V	± 0.0005%	0.00%	± 0.05%	± 0.025%	Pass	DC, FS = 60 V
AN1-Earth High Range	-10.004 V	-10.005 V	± 0.0002%	0.00%	± 0.05%	± 0.025%	Pass	DC, FS = 60 V
AN1-Earth High Range	10.034 V	10.035 V	± 0.0002%	0.00%	± 0.05%	± 0.025%	Pass	DC, FS = 60 V
AN1-Earth High Range	70.126 V	70.136 V	± 0.0005%	-0.01%	± 0.05%	± 0.025%	Pass	DC, FS = 60 V

(Certificate No. P3001143-05072015 continued on next page)

Parameter	PQube 3 Reading	Reference Reading	Reference Uncertainty X% rdg + X% FS	PQube 3 Error X% rdg + X% FS	Published Specification X% rdg + X% FS	Factory Pass/Fail X% rdg + X% FS	Pass/Fail	Notes
L8 Current 3.33V Range	1.6882 V _{rms}	1.6883 V _{rms}	± 0.006%	0.00%	± 0.05%	± 0.025%	Pass	60 Hz, FS = 3.33 V _{rms}
L8 Current 3.33V Range	4.0009 V _{rms}	4.0011 V _{rms}	± 0.007%	0.00%	± 0.05%	± 0.025%	Pass	60 Hz, FS = 3.33 V _{rms}
L1 Current 333mV Range	0.01686 V _{rms}	0.01685 V _{rms}	± 0.001%	0.00%	± 0.05%	± 0.025%	Pass	50 Hz, FS = 0.333 V _{rms}
L1 Current 333mV Range	0.16884 V _{rms}	0.16883 V _{rms}	± 0.006%	0.00%	± 0.05%	± 0.025%	Pass	50 Hz, FS = 0.333 V _{rms}
L1 Current 333mV Range	0.40016 V _{rms}	0.40017 V _{rms}	± 0.007%	0.00%	± 0.05%	± 0.025%	Pass	50 Hz, FS = 0.333 V _{rms}
L2 Current 333mV Range	0.01686 V _{rms}	0.01685 V _{rms}	± 0.001%	0.00%	± 0.05%	± 0.025%	Pass	50 Hz, FS = 0.333 V _{rms}
L2 Current 333mV Range	0.16884 V _{rms}	0.16883 V _{rms}	± 0.006%	0.00%	± 0.05%	± 0.025%	Pass	50 Hz, FS = 0.333 V _{rms}
L2 Current 333mV Range	0.40016 V _{rms}	0.40017 V _{rms}	± 0.007%	0.00%	± 0.05%	± 0.025%	Pass	50 Hz, FS = 0.333 V _{rms}
L3 Current 333mV Range	0.01686 V _{rms}	0.01685 V _{rms}	± 0.001%	0.00%	± 0.05%	± 0.025%	Pass	50 Hz, FS = 0.333 V _{rms}
L3 Current 333mV Range	0.16884 V _{rms}	0.16883 V _{rms}	± 0.006%	0.00%	± 0.05%	± 0.025%	Pass	50 Hz, FS = 0.333 V _{rms}
L3 Current 333mV Range	0.40016 V _{rms}	0.40017 V _{rms}	± 0.007%	0.00%	± 0.05%	± 0.025%	Pass	50 Hz, FS = 0.333 V _{rms}
L4 Current 333mV Range	0.01686 V _{rms}	0.01685 V _{rms}	± 0.001%	0.00%	± 0.05%	± 0.025%	Pass	50 Hz, FS = 0.333 V _{rms}
L4 Current 333mV Range	0.16884 V _{rms}	0.16883 V _{rms}	± 0.006%	0.00%	± 0.05%	± 0.025%	Pass	50 Hz, FS = 0.333 V _{rms}
L4 Current 333mV Range	0.40016 V _{rms}	0.40017 V _{rms}	± 0.007%	0.00%	± 0.05%	± 0.025%	Pass	50 Hz, FS = 0.333 V _{rms}
L5 Current 333mV Range	0.01686 V _{rms}	0.01685 V _{rms}	± 0.001%	0.00%	± 0.05%	± 0.025%	Pass	50 Hz, FS = 0.333 V _{rms}
L5 Current 333mV Range	0.16884 V _{rms}	0.16883 V _{rms}	± 0.006%	0.00%	± 0.05%	± 0.025%	Pass	50 Hz, FS = 0.333 V _{rms}
L5 Current 333mV Range	0.40016 V _{rms}	0.40017 V _{rms}	± 0.007%	0.00%	± 0.05%	± 0.025%	Pass	50 Hz, FS = 0.333 V _{rms}
L6 Current 333mV Range	0.01686 V _{rms}	0.01685 V _{rms}	± 0.001%	0.00%	± 0.05%	± 0.025%	Pass	50 Hz, FS = 0.333 V _{rms}
L6 Current 333mV Range	0.16884 V _{rms}	0.16883 V _{rms}	± 0.006%	0.00%	± 0.05%	± 0.025%	Pass	50 Hz, FS = 0.333 V _{rms}
L6 Current 333mV Range	0.40020 V _{rms}	0.40017 V _{rms}	± 0.007%	0.00%	± 0.05%	± 0.025%	Pass	50 Hz, FS = 0.333 V _{rms}
L7 Current 333mV Range	0.01686 V _{rms}	0.01685 V _{rms}	± 0.001%	0.00%	± 0.05%	± 0.025%	Pass	50 Hz, FS = 0.333 V _{rms}
L7 Current 333mV Range	0.16884 V _{rms}	0.16883 V _{rms}	± 0.006%	0.00%	± 0.05%	± 0.025%	Pass	50 Hz, FS = 0.333 V _{rms}
L7 Current 333mV Range	0.40016 V _{rms}	0.40017 V _{rms}	± 0.007%	0.00%	± 0.05%	± 0.025%	Pass	50 Hz, FS = 0.333 V _{rms}
L8 Current 333mV Range	0.01686 V _{rms}	0.01685 V _{rms}	± 0.001%	0.00%	± 0.05%	± 0.025%	Pass	50 Hz, FS = 0.333 V _{rms}
L8 Current 333mV Range	0.16884 V _{rms}	0.16883 V _{rms}	± 0.006%	0.00%	± 0.05%	± 0.025%	Pass	50 Hz, FS = 0.333 V _{rms}
L8 Current 333mV Range	0.40016 V _{rms}	0.40017 V _{rms}	± 0.007%	0.00%	± 0.05%	± 0.025%	Pass	50 Hz, FS = 0.333 V _{rms}
L1 Current 333mV Range	0.01686 V _{rms}	0.01686 V _{rms}	± 0.001%	0.00%	± 0.05%	± 0.025%	Pass	60 Hz, FS = 0.333 V _{rms}
L1 Current 333mV Range	0.16884 V _{rms}	0.16883 V _{rms}	± 0.006%	0.00%	± 0.05%	± 0.025%	Pass	60 Hz, FS = 0.333 V _{rms}
L1 Current 333mV Range	0.40016 V _{rms}	0.40017 V _{rms}	± 0.007%	0.00%	± 0.05%	± 0.025%	Pass	60 Hz, FS = 0.333 V _{rms}
L2 Current 333mV Range	0.01686 V _{rms}	0.01686 V _{rms}	± 0.001%	0.00%	± 0.05%	± 0.025%	Pass	60 Hz, FS = 0.333 V _{rms}
L2 Current 333mV Range	0.16880 V _{rms}	0.16883 V _{rms}	± 0.006%	-0.01%	± 0.05%	± 0.025%	Pass	60 Hz, FS = 0.333 V _{rms}
L2 Current 333mV Range	0.40012 V _{rms}	0.40017 V _{rms}	± 0.007%	-0.01%	± 0.05%	± 0.025%	Pass	60 Hz, FS = 0.333 V _{rms}
L3 Current 333mV Range	0.01686 V _{rms}	0.01686 V _{rms}	± 0.001%	0.00%	± 0.05%	± 0.025%	Pass	60 Hz, FS = 0.333 V _{rms}
L3 Current 333mV Range	0.16884 V _{rms}	0.16883 V _{rms}	± 0.006%	0.00%	± 0.05%	± 0.025%	Pass	60 Hz, FS = 0.333 V _{rms}
L3 Current 333mV Range	0.40016 V _{rms}	0.40017 V _{rms}	± 0.007%	0.00%	± 0.05%	± 0.025%	Pass	60 Hz, FS = 0.333 V _{rms}
L4 Current 333mV Range	0.01686 V _{rms}	0.01686 V _{rms}	± 0.001%	0.00%	± 0.05%	± 0.025%	Pass	60 Hz, FS = 0.333 V _{rms}
L4 Current 333mV Range	0.16884 V _{rms}	0.16883 V _{rms}	± 0.006%	0.00%	± 0.05%	± 0.025%	Pass	60 Hz, FS = 0.333 V _{rms}
L4 Current 333mV Range	0.40012 V _{rms}	0.40017 V _{rms}	± 0.007%	-0.01%	± 0.05%	± 0.025%	Pass	60 Hz, FS = 0.333 V _{rms}
L5 Current 333mV Range	0.01686 V _{rms}	0.01686 V _{rms}	± 0.001%	0.00%	± 0.05%	± 0.025%	Pass	60 Hz, FS = 0.333 V _{rms}
L5 Current 333mV Range	0.16884 V _{rms}	0.16883 V _{rms}	± 0.006%	0.00%	± 0.05%	± 0.025%	Pass	60 Hz, FS = 0.333 V _{rms}
L5 Current 333mV Range	0.40016 V _{rms}	0.40017 V _{rms}	± 0.007%	0.00%	± 0.05%	± 0.025%	Pass	60 Hz, FS = 0.333 V _{rms}
L6 Current 333mV Range	0.01686 V _{rms}	0.01686 V _{rms}	± 0.001%	0.00%	± 0.05%	± 0.025%	Pass	60 Hz, FS = 0.333 V _{rms}
L6 Current 333mV Range	0.16884 V _{rms}	0.16883 V _{rms}	± 0.006%	0.00%	± 0.05%	± 0.025%	Pass	60 Hz, FS = 0.333 V _{rms}
L6 Current 333mV Range	0.40016 V _{rms}	0.40017 V _{rms}	± 0.007%	0.00%	± 0.05%	± 0.025%	Pass	60 Hz, FS = 0.333 V _{rms}
L7 Current 333mV Range	0.01686 V _{rms}	0.01686 V _{rms}	± 0.001%	0.00%	± 0.05%	± 0.025%	Pass	60 Hz, FS = 0.333 V _{rms}
L7 Current 333mV Range	0.16884 V _{rms}	0.16883 V _{rms}	± 0.006%	0.00%	± 0.05%	± 0.025%	Pass	60 Hz, FS = 0.333 V _{rms}
L7 Current 333mV Range	0.40016 V _{rms}	0.40017 V _{rms}	± 0.007%	0.00%	± 0.05%	± 0.025%	Pass	60 Hz, FS = 0.333 V _{rms}
L8 Current 333mV Range	0.01686 V _{rms}	0.01686 V _{rms}	± 0.001%	0.00%	± 0.05%	± 0.025%	Pass	60 Hz, FS = 0.333 V _{rms}
L8 Current 333mV Range	0.16884 V _{rms}	0.16883 V _{rms}	± 0.006%	0.00%	± 0.05%	± 0.025%	Pass	60 Hz, FS = 0.333 V _{rms}
L8 Current 333mV Range	0.40016 V _{rms}	0.40017 V _{rms}	± 0.007%	0.00%	± 0.05%	± 0.025%	Pass	60 Hz, FS = 0.333 V _{rms}

As of this date, PQube 3 S/N P3001143 meets all factory calibration requirements, and its measurements are traceable through an unbroken chain of certified measurement standards to the United States National Institute of Standards and Technology (NIST) or other National Measurement Institutes (NMIs), or through the use of natural physical constants, intrinsic standards, or ratio calibration techniques.

For Power Standards Lab:



Matthew Muh, Senior Engineer
07 May 2015