

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST
CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE
CERTIFICATS D'ESSAIS DES EQUIPEMENTS
ELECTRIQUES (IECEE) METHODE OC

CB TEST CERTIFICATE**CERTIFICAT D'ESSAI OC**

Product
Produit

Power Quality and Energy Analyzer

Name and address of the applicant
Nom et adresse du demandeur

POWER STANDARDS LAB INC
2020 CHALLENGER DR
SUITE 100
ALAMEDA CA 94501
UNITED STATES

Name and address of the manufacturer
Nom et adresse du fabricant

POWER STANDARDS LAB INC
2020 CHALLENGER DR
SUITE 100
ALAMEDA CA 94501
UNITED STATES

Name and address of the factory
Nom et adresse de l'usine

POWER STANDARDS LAB INC
2020 CHALLENGER DR
SUITE 100
ALAMEDA CA 94501
UNITED STATES

Note: When more than one factory, please report on page 2
Note: Lorsque il y plus d'une usine, veuillez utiliser la 2^{eme} page

Ratings and principal characteristics
Valeurs nominales et caractéristiques principales

Additional Information on page 2
See Page 2

Trademark (if any)
Marque de fabrique (si elle existe)
Type of Manufacturer's Testing Laboratories used
Type de programme du laboratoire d'essais
constructeur

PSL

Model / Type Ref.
Ref. De type

PQube 3

Additional information (if necessary may also be
reported on page 2)
Les informations complémentaires (si nécessaire,,
peuvent être indiqués sur la 2^{eme} page

Additionally evaluated to EN 61010-1:2010; National Differences
specified in the CB Test Report
 Additional Information on page 2

A sample of the product was tested and found
to be in conformity with
Un échantillon de ce produit a été essayé et a été
considéré conforme à la

IEC 61010-1(ed.3), IEC 61010-2-030(ed.1)

As shown in the Test Report Ref. No. which forms part
of this Certificate
Comme indiqué dans le Rapport d'essais numéro de
référence qui constitue partie de ce Certificat

E220936-A2-CB-1 issued on 2014-03-26

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/nbcnames

Date: 2014-03-26

Signature:

Jolanta M. Wroblewska



Ref. Certif. No.

US-23086-UL

Ratings:

PQube module (ratings optionally marked):

24 Vac $\pm 10\%$, 50/60/400 Hz, 1.5 A

24-48 Vdc $\pm 10\%$, 1 A

37-57 Vdc, 15 W max (POE)

Optional PM1 plug-in module (marking required):

Input: 100-240 Vac, 50/60 Hz, 400 mA,

Aux DC Output (optionally marked): 24 Vdc, 5.15 W max

See Test Report for measurement and load ratings (optionally marked)

Additional information (if necessary)

Information complémentaire (si nécessaire)



UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA

UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK

UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN

UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2014-03-26

Signature:

Jolanta M. Wroblewska